



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/723,847	11/26/2003	Yem Chin	BSME120571	9687
26389	7590	01/10/2007	EXAMINER	
CHRISTENSEN, O'CONNOR, JOHNSON, KINDNESS, PLLC			ROZANSKI, MICHAEL T	
1420 FIFTH AVENUE			ART UNIT	PAPER NUMBER
SUITE 2800				
SEATTLE, WA 98101-2347			3768	
SHORTENED STATUTORY PERIOD OF RESPONSE		MAIL DATE	DELIVERY MODE	
3 MONTHS		01/10/2007	PAPER	

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary	Application No.	Applicant(s)
	10/723,847 Michael Rozanski	CHIN ET AL. Art Unit 3768

— The MAILING DATE of this communication appears on the cover sheet with the correspondence address —
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) Responsive to communication(s) filed on 26 November 2003.
- 2a) This action is FINAL. 2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) Claim(s) 1-11 is/are pending in the application.
 - 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) Claim(s) _____ is/are allowed.
- 6) Claim(s) 1-11 is/are rejected.
- 7) Claim(s) _____ is/are objected to.
- 8) Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on 26 November 2003 is/are: a) accepted or b) objected to by the Examiner.

Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).

Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) All b) Some * c) None of:
 1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date: _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date <u>3/13/06, 6/9/05, 2/25/04, 1/25/05</u> . | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Objections

1. Claims 1, 5, and 9 are objected to because of the following informalities:
-Claim 1 recites the limitation "the number of fringes" and "the distance" in last paragraph. Claim 5 recites the limitation "the distal end" in the second line. Claim 9 recites the limitation "the depth of insertion" in both the fourth and tenth lines. There is insufficient antecedent basis for these limitations in the claims. Appropriate action is required.

Claim Rejections - 35 USC § 103

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

3. Claims 1-5 are rejected under 35 U.S.C. 103(a) as being unpatentable over Tearney et al. in view of Bowker. In reference to claim 1, Tearney et al. disclose an imaging system comprising an endoscopic unit and an interferometer for performing multi-dimensional scanning of a structure by utilizing an optical coherence tomography technique (col. 2, lines 5-7). Although the term endoscope is used when describing the capabilities of the Tearney et al. imaging system, it is noted that the invention directly relates to guidewires, catheters, and imaging with probes placed through trocars (col. 3,

lines 20-22). Therefore, the system is capable of measuring dimensions of a body cavity. The imaging system includes an optical radiation source 2, a measuring arm 10 including an endoscopic unit 34 for being insertable into a patient's body, a reference arm 8 including a reference reflector 12 at an end thereof, and a beam divider 6 which divides light from the light source into paths defining a measuring arm and a reference arm and combines reflected radiation from the reference reflector and the structure 14 (col. 4, lines 26-29 & 44-46). As shown in Figure 4, the measuring arm 110 includes the optical fiber 32, which is capable of directing light from the source against a wall of the body cavity and receiving reflected light, and the reference arm 188 including the optical fiber 22. Figure 15 shows two detectors 417, 415 positioned in quadrature that serve to generate electrical signals representative of the interference pattern (i.e. fringes) of the combined light. The detectors transmit signals to a processing unit 418 that includes signal processing and control electronics. The magnitude of the interference pattern (i.e. fringes) is proportional to the structure's reflection coefficient the frequency of which is proportional to the differential optical path length (col. 19, lines 5-8).

Tearney et al. do not, however, specifically disclose a source of coherent light or a processor that counts fringes. In the same field of endeavor, Bowker teaches of an interferometer distance-measuring engine. The interferometer utilizes a laser light beam to provide a coherent source of light (col. 4, lines 67-68). Referring to Figure 4, one detector is in phase quadrature with a second detector, both of which are coupled to a fringe counter 69 to provide an indication of the degree and direction of displacement (col. 10, lines 71-75). Therefore, it would have been obvious to one of

Art Unit: 3768

ordinary skill in the art at the time the invention was made to incorporate a coherent light source and a fringe counter, wherein the number of fringes is proportional to the distance between an end of the optical fiber of the measuring arm and the wall of the body cavity. Coherent light is utilized in order to permit interference in regard to light waves.

In reference to claim 2, Bowker does not disclose a mechanism for rotating light emitted from optical fibers in the patient leg in the body cavity. However, Tearney et al. describes a rotational scanning mechanism 35 that causes rotation of the optical fiber 44 of the endoscopic unit 34 in the measuring arm. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to have a system further comprising a mechanism for rotating the light emitted by the optical fiber in the measuring arm within the body cavity. This enables a circumferential scan of a body cavity improved diagnosis procedures.

In reference to claim 3, Bowker does not disclose a mechanism for rotating the light including a rotatable optical coupler. However, Tearney et al. teaches of the rotational scanning mechanism 35 typically including a rotation mechanism 52 and an optical coupling system 53 (see Fig. 8). The coupling system 53 includes a coupling member 70 that is spaced by an interface 72 from an optical connector 48 affixed to the proximal end of the endoscopic unit 34. The interface 72 serves to transmit optical radiation from the input optical fiber 32 to the optical fiber 44 of the endoscopic unit (col. 9, lines 25-35). Upon activation of the drive motor 74, the shaft 78 rotates causing the rotatable optical fiber 44 to rotate (col. 9, lines 55-57). Therefore, it would have been

Art Unit: 3768

obvious to one of ordinary skill in the art at the time the invention was made to incorporate a rotatable optical coupler for coupling light to the optical fiber and a motor for rotating the optical fiber. This improves the circumferential scanning of a body cavity by incorporating a more stable rotating mechanism.

In reference to claim 4, Bowker does not disclose optical fibers routed in a catheter or a mechanism for rotating the catheter. However, Tearney et al. teach of the endoscopic unit 34 coupled to a rotational scanning mechanism 35. Tearney et al. notes that although the term endoscope is used, the invention directly relates to catheters (col. 3, lines 19-21). Since the mechanism 35 is coupled to the endoscopic unit 34, then the endoscope will rotate along with the optical fiber 44 (see Fig. 6). Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to include a mechanism for rotating a catheter with an optical fibers routed therein. This helps in transmitting and receiving the interferometric signal while within a body cavity.

In reference to claim 5, Bowker does not disclose a movable light deflector for directing emitted light in the body cavity. However, Tearney et al. describe a beam director that controls the focusing parameters of the optical beam. Further, the beam director is located on the distal end 47 of the endoscopic unit 34. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to include a movable light deflector, or beam director, at the distal end of the measuring arm to direct light in the cavity. This enables one to change the direction of the emitted light without removing the endoscope.

Art Unit: 3768

4. Claims 6-11 are rejected under 35 U.S.C. 103(a) as being unpatentable over Tearney et al. in view of Bowker, and further in view of Woker et al. In reference to claims 6, 7, 8, and 10, Tearney et al. in view of Bowker do not disclose a catheter that includes length indications or a computer that receives depth information. Woker et al. teach of indicia on the inner catheter and the instrument to indicate at least one longitudinal position of the instrument relative to the distal end of the everting element (col. 2, lines 3-6). Further, the markings may be provided in any form which is human or machine readable (col. 2, lines 42-44). In general, "machine readable" may be interpreted to include a sensor that is connected to a computer system. Tearney et al. describe a gearing mechanism made to allow a motor to axially drive the entire endoscopic imaging unit (and optical fiber) in the axial direction, thus creating three dimensional maps of the sample of interest (col. 10, lines 62-67). The imaging information from the endoscopic unit is transmitted to the signal processing, electronics control, and display unit to construct a model of the body cavity (see Fig. 1). This unit 18 functions as a computer system. Therefore, it would have been obvious to one of ordinary skill in the art at the time the invention was made to incorporate the teachings of Woker et al. to have a catheter with length indicia and a computer system that receives information regarding a depth of insertion of the catheter in order to construct a model of the body cavity.

In reference to claim 9, Tearney et al. disclose an endoscopic imaging system that is capable of measuring an internal body cavity of a patient. The system utilizes an interferometer 4 that directs a beam of light into a reference arm 8 and a measuring arm

10 that is insertable into the body cavity. A rotational scanning mechanism 35 is coupled to the endoscopic unit 34 and causes rotation of the optical fiber 44. This mechanism rotates the beam of light in the measuring arm within the body cavity and receives light via that is reflected from a wall of the body cavity via the optical fiber. The system also includes a detector 16 that detects the interference pattern and generates electrical signals representative of the combined radiation from the reference and measuring arms (col. 4, lines 47-52). The detector, therefore, is capable of detecting the difference in optical path length between the light directed into the measuring and reference arms. Tearney et al. further describe a gearing mechanism made to allow a motor to axially drive the entire endoscopic imaging unit (and optical fiber) in the axial direction, thus creating three dimensional maps of the sample of interest (col. 10, lines 62-67). The imaging information from the endoscopic unit is detected by the detector and then transmitted to the signal processing, electronics control, and display unit to construct a model of the body cavity (see Fig. 1). This unit 18 functions as a computer system.

Tearney et al., however, do not specifically disclose a coherent light source or a patient leg with a marking indicating the depth of insertion. In the same field of endeavor, Woker et al. teach of indicia on a catheter to indicate longitudinal position of the instrument (col. 2, lines 3-5). The indicia may be provided in any form which is either human or machine readable (col. 2, lines 42-44). Also in the same field of endeavor, Bowker teaches of a laser light beam that provides a coherent source of light (col. 4, lines 67-68). Therefore, it would have been obvious to one of ordinary skill in

the art at the time the invention was made to incorporate the teaches of Woker et al. and Bowker to have a system comprising an interferometer utilizing a coherent light source and a marking on the patient leg to indicate depth of insertion, a rotating mechanism, a detector to detect the optical path length difference, a computer system that receives signals from the detector, and an indication of the depth of insertion in order to construct a three-dimensional model of the body cavity.

In reference claim 11, Woker et al. do not disclose a detector including a first and second sensor positioned in quadrature. However, attention is drawn to both Tearney et al. and Bowker. Tearney et al. teach of a receiver 416 that employs two detectors 417, 415. In this case, the receiver may function as a detector unit that includes the detectors, which may function as a first and second sensor positioned in quadrature (see Fig. 15). Similarly, Bowker teaches of two detectors leading to a pair of photomultiplier tubes 65, 66 with one in phase quadrature with the other (see Fig. 4). In this case, the photomultipliers are rigidly coupled together along with other elements (col. 11, lines 16-19). Thus, the system could be described as comprising a detector unit including two photomultipliers that function as sensors positioned in quadrature. Tearney et al. describes a gearing mechanism made to allow a motor to axially drive the entire endoscopic imaging unit (and optical fiber) in the axial direction; thus creating three dimensional maps of the sample of interest (col. 10, lines 62-67). The imaging information from the endoscopic unit is transmitted to the signal processing, electronics control, and display unit to construct a model of the body cavity (see Fig. 1). This unit 18 functions as a computer system. Therefore, it would have been obvious to one of

ordinary skill in the art at the time the invention was made to include a detector that includes a first and second sensor positioned in quadrature for detecting the signals. This improves the detection of the interferometric signal by comparing the differences between the two signals.

Conclusion

5. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. The balance of art is cited to show interferometric systems and catheter markers.

US Patent No. 6,384,915 to Everett et al. which disclose an optic coupler 14, reference arm 18, reference mirror 26, sample arm 42, optical fibers, and balanced detectors for measuring body cavities.

US Patent No. 5,218,419 to Lipson et al. which disclose a fiberoptic interferometric sensor system 10 with a reference fiber 26 and a sensing fiber 24.

US Patent No. 6,036,682 to Lange et al., which disclose a catheter 30 with radiopaque segments that serve as marker bands and are visible under fluoroscopy.

6. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Rozanski whose telephone number is 571-272-1648. The examiner can normally be reached on Monday - Friday, 8-4:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eleni Mantis-Mercader can be reached on 571-272-4740. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MR

MR

Eleni Mantis Mercader
ELENI MANTIS MERCADER
SUPERVISORY PATENT EXAMINER